
smartGAS.



- Multigas analyzer
- Easy operation via touch screen
- Easy sensor calibration
- Best measurement performance
- For installation in a 19" rack
- Analog or digital output
- Photoacoustic sensors
- NDIR sensors
- Paramagnetic sensors Technology
- Electro chemical sensors



The ANAREX gas analyzers of smartGAS are high-performance solutions for various sectors of industrial gas measurement technology. The fields of application range from process measurement technology and emissions to fruit ripening. It convinces with its highly accurate and stable measurement performance as well as its customer-friendly interfacing via touch screen along with a simple sensor calibration.

The ANAREX is designed as a multi-gas analyzer and is suitable for installation in a 19" rack.

The smartGAS analyzer can also use different technologies to fulfil all kinds of application requirement.

The smartGAS Analyzer is equipped with a 5.6" touch screen, that displays the gas concentration as real-time reading. The internal software is designed to control zero, span and third point calibration to ensure a perfect accuracy of the measurement values.

Application examples

Emission monitoring CEMS Process control

Measuring principles

Infrared NDIR Electrochemical Paramagnetic O₂ **Peripherals**

Gas cooler Pumps Particle filter Tubes Mounting equipment **Customizing options**

Measuring gases Detection ranges Design and Software Gas pre treatment



General features

Measurement principle:	CO ₂ NDIR
	O ₂ electrochemical or paramagnetic
Measurement range:	0 100 Vol% (CO ₂)
	0 25 Vol% (O ₂)
Detection Limits:	≤ 1.1 Vol% (CO ₂)
	≤ 1.25 Vol% (EC) / ≤ 0.06 Vol% (PM)
Repeatabiliy	$\leq \pm 0.6 \text{ Vol}\% \text{ (CO}_2)$
	≤ 0.1 Vol% (EC) / ≤ 0.03 Vol% (PM)
Drift:	\leq ± 1.5 Vol% over 1000 h period (CO ₂)
	\leq 0.5 Vol% (EC) / \leq 0.03 Vol% (PM) over 168 h period

Basic information

Display:	Touch screen, 5.6" LCD
Analogue output:	4-20 mA (in operation) 2 mA (Warm-up or alarm)
Load resistance:	250 Ω ~ 350 Ω
Alarm output:	Device status alarm // Measurement channel level alarm
	Relay will close and isolate if alarm limits are exceeded
Relay:	1 A Trigger (250 V AC / 2 A, resistive load)
Communication:	RS232 (DB-9 Female)
Power:	(198 ~ 242) V AC, 50 / 60 Hz
Power connection:	EN 60320 C1
Fuse:	Rated current: 3A; Size: 5 x 20 mm
Protection level:	IP42 (EN 60529)
Weight:	15 kg



Gas flow & Environment information

Gas Inlet flow:	(0.4 ~ 0.8) L/min
	(flow fluctuation ≤ 0.02 L/min)
Gas temperature:	(5 ~ 35) °C
Gas pressure:	(76 ~ 116) kPa
Humidity:	non-condensing
	Inlet dew point: 5°C ± 0.1°C
Particulates:	100 μg / m³, ≤1 μm
Zero gas:	99.999 % N ₂ (NDIR)
	Clean air, free of sample gas (Photoacoustic)
Span gas:	75 % ~ 110 % of span point
	Clean air (0 \sim 25 % O ₂ measurement)
Third point gas:	35 ~ 75 % of span point
Operation temperature:	10°C ~ 40°C
Humidity:	0 ~ 95 % RH (non-condensing)
Pressure:	76 kPa ~ 116 kPa

^{*} Typical values related to 1013 hPa, Ta = 22 °C, flow = 0.7 l / min for dry (not condensing) and clean sample gas. Stated values exclude calibration gas tolerance.

All rights reserved. Any logos and/or product names are trademarks of smartGAS. The reproduction, transfer, distribution or storage of information contained in this brochure in any form without the prior written consent of smartGAS is strictly prohibited. All specifications – technical included – are subject to change without notice. Depending on the application, the target gas and the measurement range the technical data may differ. No liability is accepted for any consequential losses, injury or damage resulting from the use of this document or from any omissions or errors herein. The data is given for guidance only. It does not constitute a specification or an offer for sale.

For more information, please visit www.smartgas.eu or contact us at sales@smartgas.eu $\,$

Please consult smartGAS sales for parts specified with other temperature and measurement ranges. At first initiation and depending on application and ambient conditions recalibration is recommended. Recurring cycles of recalibration are recommended.